



Bal Seal® spring-energized seals in dental hand drills

PN-320-1 (Rev. 01, 10-31-18)

19650 Pauling
Foothill Ranch, CA
USA 92610-2610
t +1 949 460 2100
f +1 949 460 2300

VIDA Building, 1st Floor
Kabelweg 57
1014 BA Amsterdam
The Netherlands
t +31 20 638 6523
f +31 20 625 6018

Suite 901, Chinachem
Century Tower
178 Gloucester Road,
Wanchai, Hong Kong
t +852 28681860
f +852 22956753

www.balseal.com



Overview

To meet the ergonomic and surgical requirements of their operating environment, dental drill designs must be light and pencil-like. These characteristics provide users with more precise control during oral procedures.

In most drills, a rotary tool or bit is mounted at the drill head. The tool and drill head are exposed to oral surgical conditions (i.e. body fluids, temperatures) and harsh sterilization processes. Through this area of connection, the bearings and the drive unit are also exposed to the environment. These components must operate freely, and must not affect power loss or heat build-up, which can make the unit uncomfortable to hold.



Operating Parameters

Pressure

Atmospheric to autoclaved 30 psi (2 kg/cm²)

Speed

250 rpm to 30,000 rpm

Temperature

70 °F to autoclaved 275 °F (21 °C to 135 °C)

Media

Air, bearing grease, oral and sterilization fluids

Additional


Low friction and autoclavable

Features

- Bal Seal® Series KC31X, with press-in metal locking ring
- Low friction seal design minimizes heat buildup
- Bal Spring® canted coil spring energizer provides near-constant force for even wear, longer service life

Benefits

- Excellent sealing performance through autoclave sterilization cycles
- Superior protection against leakage/weepage of bearing grease from tool, and from ingress of contamination that could compromise tool performance



Air-Operated Dental Hand Drill

The information, descriptions, recommendations and opinions set forth herein are offered solely for your consideration, inquiry, and verification and are not, in part or in whole, to be construed as constituting a warranty, expressed or implied, nor shall they form or be a part of the basis of any bargain with Bal Seal Engineering, Inc.. If any sample or model was shown to or provided by Buyer/User, such sample or model was used merely to illustrate the general description and type of goods. Such use is not to be construed as a warranty that the goods will conform to the sample or model. Furthermore, THE IMPLIED WARRANTY OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE AND ALL OTHER WARRANTIES, IMPLIED OR EXPRESSED, ARE EXCLUDED AND SHALL NOT APPLY. This document provides product options for further investigation by Buyers/Users having technical expertise. The Buyer/User, through its own analysis and testing, is solely responsible for making the final selection of the products and for assuming that all performance, safety and warning requirements for the application are met. It is recommended that Buyers/Users run evaluation testing under actual service conditions to determine whether proposed Bal Seal Engineering products are suitable for the intended purpose. Nothing contained herein or in any of our literature shall be considered a license or recommendation for any use that may infringe patent rights. (LE-17)

©Copyright 2018 Bal Seal Engineering, Inc. U.S.A.